


For short implant placement, implant major diameter needs to be \leq the bur (average diameter) at the 8mm laser mark. Please refer to page 16 in the Instructions for Use Manual.

In Ridge Expansion cases, please oversize your osteotomy and make sure that the crest diameter is equal to or larger than the implant major diameter.

In Hard Bone (Mandible), after Finishing the Full Osteotomy Preparation, Use the Next Larger Size Densah™ Bur to the 3mm Laser-Mark Depth to make sure the Osteotomy Crestal Diameter is Equal to or Larger than the Implant Major (Crestal) Diameter.

Use Densah™ Burs in full-step increments for Sinus Lift cases. Example: 2.0mm, 3.0mm, 4.0mm, 5.0mm



Use large block display to compare Bur identification system when using the schematic below for proper Bur usage

● VT5® Set ○ VT8® Set ● VS8® Set

| Densifying Mode CCW (800-1500) RPMs / Cutting Mode CW (800-1500) RPMs | | | | | | | | | | | | | | | | | |
|---|---------|---------|--------------|--------------|---------------|---------------|---------------|---------------------------|---|--------------|---------------|---------------|---------------|---------------|--------------|---------------|---------------------------|
| Argon | | | K3 Pro Rapid | | | | | | | | | | | | | | |
| | | | Soft Bone | | | | | | Hard Bone (Mandible) | | | | | | | | |
| | | | | | | | | | In densifying mode make sure your osteotomy is 1.0 mm deeper than the actual implant final length. In extreme hard bone, utilize DAC (Densify After Cut) Protocol. Find protocol in IFU. | | | | | | | | |
| Geometry | Major Ø | Minor Ø | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Densah™ Bur Block Display | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Bur 5 | Bur 6 | Bur 7 | Densah™ Bur Block Display |
| Tapered | 3.00 | x | Pilot | VT1828 (2.3) | VS2228* (2.5) | — | — | | Pilot | VT1525 (2.0) | VT1828 (2.3) | VS2228* (2.5) | — | — | — | — | |
| Tapered | 3.50 | x | Pilot | VT1525 (2.0) | VT2535* (3.0) | — | — | | Pilot | VT1525 (2.0) | VT1828 (2.3) | VT2535* (3.0) | — | — | — | — | |
| Tapered | 4.00 | x | Pilot | VT1828 (2.3) | VT2838* (3.3) | — | — | | Pilot | VT1828 (2.3) | VT2535* (3.0) | VT2838 (3.3) | VS3238* (3.5) | — | — | — | |
| Tapered | 4.50 | x | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT3545* (4.0) | — | | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT2838 (3.3) | VT3545* (4.0) | — | — | — | |
| Tapered | 5.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848* (4.3) | — | | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VS4248* (4.5) | — | — | |
| Tapered | 6.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848 (4.3) | VT4858* (5.3) | | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VT4555 (5.0) | VT4858 (5.3) | VS5258* (5.5) | |

*Denotes implant placement.


*Clinician judgement and experience should be applied in conjunction with this suggestive Implant System Drilling Protocol

For short implant placement, implant major diameter needs to be \leq the bur (average diameter) at the 8mm laser mark. Please refer to page 16 in the Instructions for Use Manual.

In Ridge Expansion cases, please oversize your osteotomy and make sure that the crest diameter is equal to or larger than the implant major diameter.

In Hard Bone (Mandible), after Finishing the Full Osteotomy Preparation, Use the Next Larger Size Densah™ Bur to the 3mm Laser-Mark Depth to make sure the Osteotomy Crestal Diameter is Equal to or Larger than the Implant Major (Crestal) Diameter.

Use Densah™ Burs in full-step increments for Sinus Lift cases. Example: 2.0mm, 3.0mm, 4.0mm, 5.0mm



Use large block display to compare Bur identification system when using the schematic below for proper Bur usage

● VT5® Set ○ VT8® Set ● VS8® Set

| Densifying Mode CCW (800-1500) RPMs / Cutting Mode CW (800-1500) RPMs | | | | | | | | | | | | | | | | | | |
|---|---------|---------|--------------|--------------|--------------|---------------|---------------|---------------|---------------------------|---|--------------|--------------|--------------|---------------|--------------|---------------|---------------|---------------------------|
| Argon | | | K3 Pro Short | | | | | | | | | | | | | | | |
| | | | Soft Bone | | | | | | | Hard Bone (Mandible) | | | | | | | | |
| | | | | | | | | | | In densifying mode make sure your osteotomy is 1.0 mm deeper than the actual implant final length. In extreme hard bone, utilize DAC (Densify After Cut) Protocol. Find protocol in IFU. | | | | | | | | |
| Geometry | Major Ø | Minor Ø | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Bur 5 | Densah™ Bur Block Display | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Bur 5 | Bur 6 | Bur 7 | Densah™ Bur Block Display |
| Tapered | 4.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VS3238* (3.5) | — | — | | Pilot | VT1828 (2.3) | VT2535 (3.0) | VT2838 (3.3) | VS3238* (3.5) | — | — | — | |
| Tapered | 4.50 | x | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT3545* (4.0) | — | — | | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT2838 (3.3) | VT3545* (4.0) | — | — | — | |
| Tapered | 5.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848 (4.3) | VS4248* (4.5) | — | | Pilot | VT1828 (2.3) | VT2535 (3.0) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VS4248* (4.5) | — | |
| Tapered | 6.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848 (4.3) | VT4858 (5.3) | VS5258* (5.5) | | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VT4555 (5.0) | VT4858 (5.3) | VS5258* (5.5) | |

*Denotes implant placement.


*Clinician judgement and experience should be applied in conjunction with this suggestive Implant System Drilling Protocol

For short implant placement, implant major diameter needs to be \leq the bur (average diameter) at the 8mm laser mark. Please refer to page 16 in the Instructions for Use Manual.

In Ridge Expansion cases, please oversize your osteotomy and make sure that the crest diameter is equal to or larger than the implant major diameter.

In Hard Bone (Mandible), after Finishing the Full Osteotomy Preparation, Use the Next Larger Size Densah™ Bur to the 3mm Laser-Mark Depth to make sure the Osteotomy Crestal Diameter is Equal to or Larger than the Implant Major (Crestal) Diameter.

Use Densah™ Burs in full-step increments for Sinus Lift cases. Example: 2.0mm, 3.0mm, 4.0mm, 5.0mm



Use large block display to compare Bur identification system when using the schematic below for proper Bur usage

● VT5® Set ○ VT8® Set ● VS8® Set

| Densifying Mode CCW (800-1500) RPMs / Cutting Mode CW (800-1500) RPMs | | | | | | | | | | | | | | | | | | |
|---|---------|---------|-------------|--------------|---------------|---------------|---------------|---------------|---------------------------|---|--------------|--------------|---------------|---------------|---------------|--------------|---------------|---------------------------|
| Argon | | | K3 Pro Sure | | | | | | | | | | | | | | | |
| | | | Soft Bone | | | | | | | Hard Bone (Mandible) | | | | | | | | |
| | | | | | | | | | | In densifying mode make sure your osteotomy is 1.0 mm deeper than the actual implant final length. In extreme hard bone, utilize DAC (Densify After Cut) Protocol. Find protocol in IFU. | | | | | | | | |
| Geometry | Major Ø | Minor Ø | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Bur 5 | Densah™ Bur Block Display | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Bur 5 | Bur 6 | Bur 7 | Densah™ Bur Block Display |
| Straight | 3.00 | x | Pilot | VT1828 (2.3) | VS2228* (2.5) | — | — | — | | Pilot | VT1525 (2.0) | VT1828 (2.3) | VS2228* (2.5) | — | — | — | — | |
| Straight | 3.50 | x | Pilot | VT1525 (2.0) | VT2535* (3.0) | — | — | — | | Pilot | VT1525 (2.0) | VT1828 (2.3) | VT2535* (3.0) | — | — | — | — | |
| Straight | 4.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VS3238 (3.5) | — | — | | Pilot | VT1828 (2.3) | VT2535 (3.0) | VT2838 (3.3) | VS3238* (3.5) | — | — | — | |
| Straight | 4.50 | x | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT3545* (4.0) | — | — | | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT2838 (3.3) | VT3545* (4.0) | — | — | — | |
| Straight | 5.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848 (4.3) | VS4248* (4.5) | — | | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VS4248* (4.5) | — | — | |
| Straight | 6.00 | x | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848 (4.3) | VT4858 (5.3) | VS5258* (5.5) | | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VT4555 (5.0) | VT4858 (5.3) | VS5258* (5.5) | |

*Denotes implant placement.

*Clinician judgement and experience should be applied in conjunction with this suggestive Implant System Drilling Protocol